



TempTale® Ultra

Serial #: JME4N0BJK0

Alarm Status

OK

Monitor Configuration

Start Up Delay: 20 min Interval: 5 min

Monitor Read

On : 16-FEB-2022 1:06:40 AM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 27-JAN-2022 06:59:42 AM
Stop Time: 15-FEB-2022 11:59:56 PM
Number of Points: 5677
Trip Length: 19 day 17 hr 5 min

Summary Data

Low Extreme: 2.8 °C @ 01-FEB-2022 2:24:42 PM
High Extreme: 7.0 °C @ 27-JAN-2022 06:59:42 AM
Mean ± Std Deviation: 4.2 °C ± 0.4 °C
Mean Kinetic Temperature: 4.2 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm Data

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
-------	-----------------	----------------	------------------	------------	------------------	-----------	--------

Originator Notes:

