



# TempTale® Ultra

Serial #: JME4N0BNT0

### Alarm Status

OK

### Monitor Configuration

Start Up Delay: 20 min      Interval: 5 min

### Monitor Read

On : 16-FEB-2022 1:08:50 AM      By : TempTale® Ultra  
Note: All Times in GMT

### Recorded Data

First Point: 27-JAN-2022 6:59:38 AM  
Stop Time: 16-FEB-2022 12:00:01 AM  
Number of Points: 5677  
Trip Length: 19 day 17 hr 5 min

### Summary Data

Low Extreme: 4.9 °C @ 14-FEB-2022 1:29:38 AM  
High Extreme: 8.8 °C @ 27-JAN-2022 6:59:38 AM  
Mean ± Std Deviation: 5.6 °C ± 0.2 °C  
Mean Kinetic Temperature: 5.6 °C  
Activation Energy: 83.144 kJ/mol

### Alarm Summary

### Alarm Data

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status

### Originator Notes:

