



TempTale® Ultra

Serial #: JME4N0BHGO

Alarm Status

OK	
----	--

Monitor Configuration

Start Up Delay:	20 min	Interval:	5 min
-----------------	--------	-----------	-------

Monitor Read

On :	16-FEB-2022 1:03:26 AM	By :	TempTale®Ultra
Note:	All Times in GMT		

Recorded Data

First Point:	27-JAN-2022 6:44:32 AM
Stop Time	15-FEB-2022 11:59:14 PM
Number of Points:	5679
Trip Length:	19 day 17 hr 15 min

Summary Data

Low Extreme:	20.4 °C @ 09-FEB-2022 7:24:32 PM
High Extreme:	21.4 °C @ 30-JAN-2022 11:39:32 PM
Mean ± Std Deviation:	20.7 °C ± 0.2 °C
Mean Kinetic Temperature:	20.7 °C
Activation Energy:	83.144 kJ/mol

Alarm Summary

Alarm Data

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status

Originator Notes:

